Phone: +33 (0) 466 717 184 Fax: +33 (0) 466 532 305

Email: export@nausicaa-medical.com

VAT N°: FR 954 208 72 293



EC DECLARATION OF CONFORMITY

In accordance with Annex of Decree No. 95-292 of the Code of Public Health (Annex VII of Directive 93/42 EEC amended by 2007/47/EEC).

We certify hereby that the products listed below comply with the requirements of the EU Directive 93/42/EEC on medical devices. This certificate loses its value in case of inappropriate use of these products without written authorization from the manufacturer. We ensure and declare that these devices belong to Class I and meet the provisions of the decree which they are subject to. Therefore, I consider that I comply with the obligations of the essential requirements of Annex I.

FLYER XL with electrical opening legs

Compliance with the following standards:

Medical products (93/42 EEC amended by 2007/47/EEC)

Electromagnetic compatibility (2004/108/EC)

Applicable standards:

• NF EN 60601-1: 2007

Safety for medical electrical equipment

• NF EN 60601-1-2: 2007

Electromagnetic compatibility

NF EN ISO 10535: 2007

Hoists for transfer of disabled persons-Requirements and test methods

Manufacturer: NAUSICAA Médical

12 allée du Piot ZA Pôle Actif

30660 GALLARGUES LE MONTUEUX

France

Gallargues-le-Montueux, February 21th, 2017

